



CERTIFICATE OF ACCREDITATION

ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

North American Science Associates, Inc. (NAMSA)

9 Morgan

Irvine, CA 92618

has been assessed by ANAB and meets the requirements of international standard

ISO/IEC 17025:2017

while demonstrating technical competence in the field of

TESTING

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AT-2562

Certificate Number


ANAB Approval

Certificate Valid Through: 03/09/2020
Version No. 005 Issued: 06/06/2019



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



ANSI National Accreditation Board

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

North American Science Associates, Inc. (NAMSA)

9 Morgan
Irvine, CA 92618

Valerie D. Gnepper 419-666-9455
vgnepper@namsa.com www.namsa.com

TESTING

Valid to: **March 9, 2020**

Certificate Number: **AT-2562**

Chemical

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
USP Physiochemical Testing – Polymers, Resins, Mixed Materials, Elastomeric Closures	USP <661>; USP <381>	Polymers, Non-Volatile Residue, Particulates	Balances, Incubators, pH Meter, UV-Vis Spectrophotometer
Infrared Analysis (IR)	ISO 10993-18; USP <197>	Polymers, Non-Volatile Residue, Particulates	FTIR Spectrophotometer
Particulate Analysis	USP <788>; USP <789>; ISO 14708-3	Medical Devices	Particle Counting Apparatus, LAF Hood, Microscope
Ethylene Oxide Residuals Analysis	ISO 10993:7	Medical Devices	Gas Chromatography
USP Total Organic Carbon	USP <643>	Medical Devices	TOC Analyzer

Biological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Cytotoxicity Assay, Elution Method (48-hour Exposure): L929 Cells	ISO 10993-5; USP <87>	Polymers, Metals, Assembled Devices, Materials	Cell Culture Equipment, Microscope
Cytotoxicity Assay, Agarose Overlay: L929 Cells	ISO 10993-5; USP <87>	Polymers, Metals, Assembled Devices, Materials	Cell Culture Equipment, Microscope
USP Limulus Amebocyte Lysate (LAL) Kinetic Turbidimetric Method	USP <85>	Polymers, Metals, Assembled Devices, Materials, Drug Products (Intermediates and Finished)	Spectrophotometer, Incubator
USP Limulus Amebocyte Lysate (LAL) Kinetic Chromogenic Method	USP <85>	Polymers, Metals, Assembled Devices, Materials, Drug Products (Intermediates and Finished)	Spectrophotometer, Incubator
USP Limulus Amebocyte Lysate (LAL) Gel Clot Method	USP <85>	Polymers, Metals, Assembled Devices, Materials, Drug Products (Intermediates and Finished)	Incubator, Qualitative Observation

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Bioburden Testing of Medical Products	ISO 11737-1; USP<55>; USP<61>; USP<1231>	Medical Devices, Materials, Pharmaceuticals, Human Tissues	ISO Class 5 Hoods, Incubators
Identification of Microorganisms	Applied Biosystems MicroSEQ® Microbial Identification System Analysis and Interpretation	Medical Devices, Materials, Pharmaceuticals, Human Tissues	ISO Class 5 Hoods, Genetic Sequencer, Thermocyclers
Bacteriostasis/Fungistasis Testing	ISO 11737-2; ISO 11137-2; ISO 11135; USP <71>	Medical Devices, Materials, Pharmaceuticals, Human Tissues	Class 6 Cleanroom, ISO Class 5 Hoods, Incubators
USP Antimicrobial Preservative Effectiveness	USP <51>	Medical Devices, Materials, Pharmaceuticals	ISO Class 5 Hoods, Incubators

Microbiological

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Bioburden Recovery Validation	ISO 11737-1; USP<1227>	Medical Devices, Materials, Pharmaceuticals, Human Tissues	ISO Class 5 Hoods, Incubators
Incubation and Enumeration of Fallout Plates, RODAC® Plates, or Air Sampler Media	ISO 14698-1; USP<1116>	Environmental Monitoring	Incubators, Macroscopic Observations
Kirby-Bauer Standard Antimicrobial Susceptibility Test	American Journal of Clinical Pathologists, Volume 95	Medical Devices, Materials, Pharmaceuticals	ISO Class 5 Hoods, Incubators
Determining Antimicrobial Activity of Immobilized Antimicrobial Agents	ASTM E2149	Medical Devices, Materials	ISO Class 5 Hoods, Incubators
Sterility	ISO 11737-2; ISO 11137-2; ISO 11135; USP <71>	Medical Devices, Materials, Pharmaceuticals, Human Tissues	ISO Class 6 Cleanroom, ISO Class 5 Hoods, Incubators
Gram Stain and Colony Morphology	FDA Microbiological Methods & Bacteriological Analytical Manual (BAM)	Medical Devices, Materials, Pharmaceuticals, Human Tissues	ISO Class 5 Hoods, Incubators, Microscopes

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-2562.



Vice President